




# Certificate

<b>Product:</b>	Seraclone Anti-D (RH1) Blend	<b>REF</b>	12013865
<b>Intended Use:</b>	Seraclone Anti-D (RH1) Blend is used for the qualitative detection of the D antigen on human red blood cells in the slide and tube test in patient and donor samples.	<b>LOT</b>	3237092-00
			2022-09-15
			2024-09-04
		<b>ACT</b>	1:32
		<b>VOL</b>	10 ml
			2°C - 8°C

- **Specifications:**

**Clone:** BS232 (IgM) / BS221 (IgG) / H41 11B7 (IgG), human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

***This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00***