




Certificate

| | | | |
|----------------------|---|---|------------|
| Product: | Seraclone Anti-D (RH1) Blend | REF | 12013865 |
| Intended Use: | Seraclone Anti-D (RH1) Blend is used for the qualitative detection of the D antigen on human red blood cells in the slide and tube test in patient and donor samples. | LOT | 3408151-00 |
| | |  | 2024-02-23 |
| | |  | 2026-02-12 |
| | | ACT | 1:32 |
| | | VOL | 10 ml |
| | |  | 2°C - 8°C |
| | | | |

- **Specifications:**

Clone: BS232 (IgM) / BS221 (IgG) / H41 11B7 (IgG), human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00