




Certificate

Product:	Seraclone Anti-D (RH1) 226	REF	12013866
Intended Use:	Seraclone Anti-D (RH1) 226 is used for the qualitative detection of the D antigen of human red blood cells in the slide and tube test in patient and donor samples.	LOT	3238141-00
			2022-09-23
			2024-09-12
		ACT	1:32
		VOL	10 ml
			
		2°C – 8°C	

- **Specifications:**

Clone: BS226 (IgM), human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00