

# Certificate

Product:	Seraclone Anti-E (RH3)	REF	12013869
Intended Use:	Seraclone Anti-C (RH2), Seraclone Anti-c (RH4), Seraclone Anti-E (RH3), Seraclone Anti-e (RH5) are used for the qualitative detection of the C (RH2), c (RH4), E (RH3) and e (RH5) antigens of human red blood cells in the slide and tube test in patient and donor samples.	LOT	3334040-00
		M	2023-08-22
		R	2024-11-14
		ACT	1:64
		VOL	5 ml
		2°C	

## • Specifications:

Clone: MS258 (IgM) / 906 (IgM), human monoclonal

### • Information, Instructions and Safety:

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

## • Analytic Results, Quality Control:

- The product lot was released following evaluation of sensitivity and specificity.

#### Regulation:

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00