




# Certificate

<b>Product:</b>	Seraclone Anti-E (RH3)	<b>REF</b>	12013869
<b>Intended Use:</b>	Seraclone Anti-C (RH2), Seraclone Anti-c (RH4), Seraclone Anti-E (RH3), Seraclone Anti-e (RH5) are used for the qualitative detection of the C (RH2), c (RH4), E (RH3) and e (RH5) antigens of human red blood cells in the slide and tube test in patient and donor samples.	<b>LOT</b>	3439020-00
			2024-09-27
			2025-12-21
		<b>ACT</b>	1:64
		<b>VOL</b>	5 ml
			2°C - 8°C

- **Specifications:**

**Clone:** MS258 (IgM) / 906 (IgM), human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

***This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00***