




Certificate

Product:	Seraclone Anti-e (RH5)	REF	12013870
Intended Use:	Seraclone Anti-C (RH2), Seraclone Anti-c (RH4), Seraclone Anti-E (RH3), Seraclone Anti-e (RH5) are used for the qualitative detection of the C (RH2), c (RH4), E (RH3) and e (RH5) antigens of human red blood cells in the slide and tube test in patient and donor samples.	LOT	3436070-00
			2024-09-05
			2026-08-26
		ACT	1:8
		VOL	5 ml
			2°C - 8°C

- **Specifications:**

Clone: MS16 (IgM) / MS21 (IgM) / MS63 (IgM), human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the Instructions for Use and the Material Safety Data Sheet of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- The product has been manufactured and tested by Bio-Rad Medical Diagnostics GmbH in accordance with the currently applicable regulations for the manufacture of in-vitro diagnostics.

This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00