



## Certificate / Zertifikat

<b>Product / Produkt:</b>	<b>Solidscreen II Anti-D (RH1) Blend</b> Solidscreen II Human Monoclonal Blend	<b>REF</b>	806530
<b>Intended Use/ Verwendungszweck:</b>	Solidscreen II Anti-D (RH1) Blend is used with Solidscreen II for the detection of weak D and partial D antigens (DVI and DVII). It is used to test samples which have been tested RhD negative with a direct agglutination method. Solidscreen II Anti-D (RH1) Blend wird im Solidscreen II-Festphasen-Test zum Nachweis des Blutgruppenmerkmals weak D und partial D (DVI und DVII) bei RhD negativen Blutproben eingesetzt.	<b>LOT</b>	3232051-00
			2024-08-10
			2 ... 8 °C

- **Specifications**

**Clone:** BS221 (IgG) / H41 11B7 (IgG) human monoclonal

- **Information, Instructions and Safety:**

- For further information, instructions and safety refer to the included instructions for use of this product.

- **Analytic Results, Quality Control:**

- The product lot was released following evaluation of sensitivity and specificity.

- **Regulation:**

- DIN EN ISO 13485
- MDSAP
- Directive 98/79/EC
- FDA 21 CFR Part 820, 809, 600, relevant only for products for the US-market
- Recommended Methods for Blood Grouping Reagents Evaluation, Docket No.84S-0181, March 1992

***This document has been produced electronically and is valid without a signature, according to SOP SV-D:A-042-00 /***

***Dieses Dokument wird maschinell erstellt und ist auch ohne Unterschrift gültig, gemäß SV-D:A-042-00***